

Cystatin C

Gentian Cystatin C Calibrator Kit

REF 1051

For *in vitro* diagnostic use by laboratory professionals.

This document describes the general use of the product above. For instrument specific settings, please refer to the application notes available upon request to marketing@gentian.com.

Intended purpose

The Gentian Cystatin C Immunoassay is an immunoturbidimetric assay intended for the *in vitro* quantitative determination of cystatin C in human serum and plasma on automated clinical analysers by laboratory professional users. The measurement of cystatin C is used in the diagnosis and treatment of renal diseases.

Calibrator kit indication for use

The Gentian Cystatin C Calibrator Kit is intended to be used to establish a calibration curve for measuring cystatin C concentration in human plasma and serum samples with the Gentian Cystatin C Immunoassay.

Calibrator value assignment

The calibrator value, given in the analytical value sheet, is assigned according to Gentian's value transfer protocol as recommended in ISO 17511 [1] for calibrators and controls. For lot specific concentrations, please consult the analytical value sheet available on www.gentian.com.

Calibrator standardisation

The Gentian Cystatin C Calibrator Kit is standardised against the international calibrator standard ERM-DA471/IFCC.

Assay kit components

Products provided	
Gentian Cystatin C Calibrator Kit (6 levels x 1 mL)	REF 1051
Products required, but not provided	
Gentian Cystatin C Control Kit (2 levels x 1 mL), or	REF 1019
Gentian Cystatin C Control Kit (2 levels x 5 mL)	REF 1026
Gentian Cystatin C Reagent Kit	REF 1101
<ul style="list-style-type: none"> R1 Assay Buffer (58 mL) R2 Immunoparticles (10 mL) 	

All products are ready for use.

Composition

The Gentian Cystatin C Calibrator Kit consists of delipidated human serum pools spiked with human cystatin C. Antibiotics are used as preservation. For lot specific concentrations, please consult the analytical value sheet available on www.gentian.com.

Warnings and precautions

1. Contains substances from human or animal origin and should be considered as potentially infectious material. Handle with caution and discard following local regulations.
2. The sodium azide concentration of the assay is not characterised as hazardous. Although, accumulated NaN_3 in lead and copper pipes may cause generation of explosive metal azides. To prevent this, rinse thoroughly if discarded into the drain.
3. Contains a sensitizing substance below concentration limit. May produce an allergic reaction in some people and may cause respiratory irritation if inhaled.

4. Contains antibiotics and must be handled with due caution.
5. Exposure may result in irritation of skin and eyes.
6. Avoid contact with incompatible materials.
7. Avoid exposure to heat and direct sunlight.

For additional safety information, please refer to the SDS (Safety Data Sheet) available on www.gentian.com.

Additional handling instructions

1. This test is for *in vitro* use only and must be handled by laboratory professionals.
2. Use only validated and approved instrument applications.
3. Do not use products after the expiration date has passed.
4. Do not mix reagents of different lots or interchange caps of reagents, controls, calibrators, and lots.
5. Tighten caps carefully back on after use of reagents, calibrators, and controls to avoid evaporation.

Reagent storage and stability

All products provided for the Gentian Cystatin C Immunoassay must be stored at 2-8 °C. The expiry date is printed on the labels. The in-use stability of the Gentian Cystatin C Calibrator Kit was found to be at least 26 weeks performed as an open vial study (at 2-8 °C) based on CLSI guideline EP25 [2]. For calibration curve stability, please refer to the instrument specific application notes.

Procedure

Application notes

Applications of the Gentian Cystatin C Immunoassay have been established on several clinical chemistry analysers. Detailed, validated application notes describing the procedures for installation and analysis on specific instruments are available upon request from marketing@gentian.com. For instructions on how to install a new application, consult the instrument manual. Maintenance, operation, and precautions must be handled in accordance with the specific instrument manual.


Establishment of the calibration curve

The calibrator levels 1 to 6 are used to establish a 6-point standard curve as defined in the instrument manual. The calibrators are ready to use, do not program dilution. Calibrator values are lot specific, and calibration must be performed whenever a new lot is used. The values assigned to the calibrator kit are provided in the analytical value sheet available on www.gentian.com. A new calibration should be performed according to the instrument specific calibration curve stability or when a new reagent lot is used.

QC controls

Gentian Cystatin C Controls must be tested each day before any samples are measured in order to validate the calibration curve. The controls have lot specific concentration ranges that must be met before measuring samples. The assigned value ranges are given in the analytical value sheet available on www.gentian.com. If the control values measured are not valid, repeat the control measurements. Recalibrate if necessary. If the calibration cannot be performed without error, or valid control values cannot be reproduced, contact the local distributor for support.

Symbols key



Temperature limit



Use by date



Consult instructions for use



Manufacturer



CE mark with Notified Body number



UKCA mark



Swiss authorized representative



In Vitro Diagnostic medical device



Lot number



Catalogue number



Unique Device Identifier



Contents



Calibrator



Calibrator level 1



Calibrator level 2



Calibrator level 3



Calibrator level 4



Calibrator level 5



Calibrator level 6



Gentian AS

Bjornasveien 5
N-1596 Moss
Norway
TEL: +47 99 33 99 05
www.gentian.com



Representatives

UK Responsible
Person

Emergo Consulting (UK) Limited
c/o Cr360 – UL International
Compass House, Vision Park Histon
Cambridge CB24 9BZ
United Kingdom



MedEnvoy Switzerland
Gotthardstrasse 28
6302 Zug
Switzerland

References

1. EN ISO 17511:2021 In vitro diagnostic medical devices — Requirements for establishing metrological traceability of values assigned to calibrators, trueness control materials and human samples. (ISO 17511:2020).
2. CLSI. *Evaluation of stability of in vitro Diagnostic Reagents; Approved Guideline*. CLSI document EP25-A. Wayne, PA: Clinical and Laboratory Standards Institute; 2009.

Serious incidents

Please notify the manufacturer and your competent authority if any serious incidents have occurred in relation to the device.

Modifications from previous version

- The availability of SDS in Gentian webpage is updated under the warnings and precautions section

Date of issue

2026-06-26